## **DEKO 190IX WASHER DISINFECTOR - LOADING INSTRUCTIONS**

(ve rsio n: 10/16)



**BOWLS** Do not empty fo a ming liquids into machine.



SLIPPER PANS, BOTILES & LIDS No pre-sluic ing require d



BEDPANS AND BOTTLES No pre-sluic ing require d



**LARGE BOWLS** Maximum 2 perload

## CYCLE INFORMATION - TO UCH SCREEN DISPLAY



ABOVE: TO UCH SCREEN USER DISPLAY PANEL

NORMAL: 10 MINUIES INCLUDING FLUSH, SHORT WASH & DISINFECTION IDEAL FOR BOWLS, BOTILES.

INTENSIVE: 11 MINUTES INCLUDING FLUSH, WASH & DISINFECTION IDEAL FOR PANS AND BOTILES.

EMPTYING: NON DISINFECTION CYCLE 1 MINUTE

(Note: This cycle may be disabled on your machine) FOR FLUSHING OUT CONTENTS OF PAN IF YOU DO NOT HAVE A FULLIOAD TO

PROCESS.

WASTE

CONFIRMATION OF NON-DISINFECTION CYCLE REQUIRED BY USER PRIOR TO START.

LONG: 13 MINUTES FLUSH, WASH, DISINFECTION. LONG EST PRE SET CYCLE WITH 5 MINUTE WASH PHASE Use ful for SOILED MOP BUCKEIS, CYTOTOXIC

## IMPORTANT:

- DO NOT PLACE ANY HEMS IN THE MACHINE THATCANNOTGO INTO A TOILET NO WIPES, PAPER TO WELL TESTING STRIPS, EIC.
- SMAIL FIEMS MUST BE PLACED IN THE UTENSIL BASKETPRO VIDED.
- IF YO U ARE NO TSURE ASK A NURSE EDUCATOR.



**SMALL ITEMS** MUSTGO IN THE UTENSIL BASKET **PRO VIDED**